ORIGINAL ARTICLE

Evaluation of the Simplify D-dimer assay as a screening test for the diagnosis of deep vein thrombosis in an emergency department

D Neale, C Tovey, A Vali, S Davies, K Myers, M Obiako, V Ramkumar, A Hafiz

Emerg Med J 2004;21:663–666. doi: 10.1136/emj.2003.011049

See end of article for authors' affiliations

Correspondence to: Mr C Tovey, Emergency Department, Prince Charles Hospital, Merthyr Tydfil, Mid Glamorgan, Wales CF47 9DT, UK; clive.tovey@nglam-tr.wales.nhs.uk

Accepted for publication 5 February 2004

Objectives: To evaluate the use in an emergency department of a new D-dimer assay (Simplify D-dimer) as a screening test for deep vein thrombosis (DVT).

Methods: 187 outpatients with clinical features suspicious of acute DVT were entered into this study. A Simplify D-dimer test was performed in the emergency department on all patients. A SimpliRED D-dimer test and a semi-automated latex agglutination assay (Auto-D-dimer 700 on a Thromboscreen 400C analyser) were performed in the haematology laboratory. All patients were investigated with contrast venography to confirm or exclude the diagnosis of DVT.

Results: The Simplify test had a sensitivity of 94.1% and a negative predictive value (NPV) of 94.8%. These results compared favourably with the SimpliRED test (sensitivity 74.5%, NPV 89.7%) and the latex agglutination assay (sensitivity 90.2%, NPV 92.2%). This increased sensitivity was at the cost of a lower specificity, the specificity of the three D-dimer tests being Simplify 40.4%, SimpliRED 83.1%, and latex agglutination 43.4%.

Conclusions: Simplify proved to be a rapid and easy to use test and may be useful for use in the emergency department as part of a diagnostic algorithm for deep vein thrombosis. Further larger scale studies are needed.

Venous thromboembolic disease has an estimated annual incidence in developed countries of 1 in 1000 people.¹ The clinical diagnosis of deep vein thrombosis (DVT) of the lower limb is unreliable and individual symptoms and signs are of little value.^{2–3}

Contrast venography has traditionally been the “gold standard” for the diagnosis of DVT⁴ but venography has been replaced in most centres by colour flow duplex ultrasonography. Venography and ultrasonography are comparatively expensive investigations and over 75% of the investigated patients will prove not to have a DVT.^{5–6} Venography is also invasive and associated with a small but significant risk of complications.⁴

A screening test is needed with a high negative predictive value (NPV), which can thus be used to reduce the need for imaging. A fast, reliable, non-invasive, and inexpensive screening test is needed that can be performed immediately on attendance at hospital. In recent years interest has focused on the use of D-dimer assays and plethysmography.⁷

D-dimers are specific cross linked derivatives of fibrin, produced when fibrin is degraded by plasmin, so concentrations are raised in patients with venous thrombosis.⁸ D-dimer tests may be negative in cases of proven acute DVT for a number of reasons:

- D-dimer assays use a monoclonal antibody against the epitope on the D-dimer fragment that is absent on fibrin, fibrinogen, and non-cross linked fragments of fibrin. The reactivity of different D-dimer monoclonal antibodies to these different molecular weight species varies, causing variable results with different assays within individual patients.^{9–10}
- In isolated calf vein DVT the thrombus load will be less than in extensive proximal DVT. Patients with isolated calf vein DVT will therefore be expected to have lower plasma

D-dimer concentrations and this has been shown in some studies.^{11–12}

- In the occasional patient there may be impaired fibrinolytic activity.¹³
- With the SimpliRED test there is a visual end point of agglutination of red blood cells and some studies have shown that interobserver variability in test results is high.^{14–15}

D-dimer tests generally have a high NPV but cannot be used in isolation as screening tests. D-dimer tests are therefore often used in conjunction with clinical probability scoring, plethysmography, or ultrasonography to reduce the need for further imaging.^{16–20} Although sensitive for venous thrombosis, high concentrations of D-dimer are insufficiently specific for making a positive diagnosis because they occur in other disorders such as malignancy, pregnancy, inflammatory conditions, and after surgery.

There are four types of D-dimer assays commercially available—enzyme linked immunosorbent assays (ELISA), latex agglutination assays, a whole blood agglutination assay (SimpliRED), and an immunochromatography test (Simplify). The SimpliRED and Simplify tests are suitable for use as bedside tests whereas the ELISA and latex agglutination tests need to be performed in a laboratory.

There are many quantitative latex agglutination and ELISA tests available^{9–10–21} and the conventional ELISA is considered to be the gold standard for determination of D-dimer concentration. However, these laboratory based methods have the disadvantage that there can be a considerable delay in receiving the results in the emergency department.

Abbreviations: DVT, deep vein thrombosis; ELISA, enzyme linked immunosorbent assay; NPV, negative predictive value; PPV, positive predictive value

Figure 1 Photograph of a positive Simplify D-dimer test with a line visible in the test zone. A colour version of the figure is available on the journal web site (<http://www.emjonline.com/supplemental>).

Moreover, the laboratory based methods are more suited to the assay of many samples in one batch rather than individual samples performed as soon as the patient attends the emergency department. A number of rapid D-dimer assays are now available and include Tinaquant, Liatest, and Vidas D-dimer. These tests have sensitivities similar to conventional ELISA tests.²¹

The SimpliRED D-dimer test has been used widely as a near patient test in emergency departments and provides a result within five minutes. However, the sensitivity of the SimpliRED test has been reported to vary from 61% to 100%.⁸

A new near patient assay (Simplify D-dimer) has recently been introduced but there are currently no publications that have evaluated its efficacy. The aim of this study was to evaluate the use in the emergency department of the Simplify D-dimer test. The Simplify D-dimer test was compared with the (near patient) SimpliRED D-dimer assay and a (laboratory based) latex agglutination assay (Auto D-dimer 700 on a Thromboscreen 400C analyser).

METHODS

The subjects were recruited into the study over a 21 month period starting in April 2001. The study was approved by the Bro-Taf Health Authority Local Research Ethics Committee and the subjects all gave informed written consent.

Patients presenting in the emergency department with clinical features suspicious of DVT were entered into the study, but the pre-test probability scores as performed in some studies¹¹⁻¹⁸ were not calculated. Contrast venography was used to investigate DVT in all patients. As contrast venography is considered to be the gold standard, there was no follow up of patients with a negative venogram.

Table 1 D-dimer test and venogram results of the 187 patients

		Venogram positive	Venogram negative
Simplify	positive	48	81
	negative	3	55
SimpliRED	positive	38	23
	negative	13	113
Auto D-dimer	positive	46	77
	negative	5	59

Patients were excluded from the study if they were less than 18 years old, had experienced recent trauma (<6 weeks), had undergone recent surgery (<6 weeks), were pregnant, had an underlying malignancy, or if they were having anticoagulant treatment. Patients were also excluded from the study if we were unable to perform venography (because of technical difficulties, or previous reaction to contrast).

Venous blood was obtained from each patient and a Simplify D-dimer test was performed immediately in the emergency department. Venous blood was also collected into plastic 4.5 ml Tri-sodium citrate (0.105M) Becton-Dickinson vacutainers in a ratio of 9:1 (blood:anticoagulant) and sent to the haematology laboratory for assessing with the SimpliRED and latex agglutination tests.

Simplify D-dimer

The Simplify D-dimer assay consists of an individual test device in which the active ingredients are murine monoclonal antibody (DD3B6/22) specific for D-dimer conjugated to colloidal gold particles, a second D-dimer specific murine monoclonal antibody, and a sheep antimurine IgG antibody. Some 35 µl of whole blood is added to the sample well followed by two drops of buffer. The device was left lying flat for 10 minutes and then checked for a pink/purple line in the procedural control zone of the reading window. The absence of a line in this zone renders the test invalid. A positive result was indicated by the presence of a pink/purple line in the test zone of the reading window indicating a D-dimer concentration of greater than 80 ng/ml. A negative result was indicated by a complete absence of a line in the test zone of the reading window.

SimpliRED D-dimer

The SimpliRED D-dimer tests were performed in the haematology laboratory on the citrated whole blood. The methodology for the SimpliRED D-dimer assay has been described elsewhere.²² The assay contains a bispecific antibody that reacts with high affinity to D-dimer (3B6/22), and with a red cell binding antibody (RAT-1C3/86). In the presence of D-dimer concentrations >120 ng/ml, the antibody induces agglutination of the patient's red blood cells.

Latex agglutination test

After centrifugation of the citrated venous blood at 2000–3000 g for 10 minutes, the plasma was then separated and frozen at –20°C for a maximum of one month. These samples were batch analysed for D-dimer using the Auto D-dimer (700) kit on the Thromboscreen 400C analyser.

The Auto D-dimer (700) is a quantitative latex micro-particle enhanced turbidimetric immunoassay that uses monoclonal antibody coated (MA-8D3) latex particles. In the presence of D-dimer, the particles aggregate thus increasing turbidity. The increase in scattered light is proportional to the amount of D-dimer in the sample.

The turbidity change in the sample is measured on a Thromboscreen 400C photometer that is set to measure at 705 nm. Photometric detection is started when the latex reagent is added to a cuvette containing the patient's sample. The change in optical density (OD = Extinction) is monitored by the instrument. The turning point of the reaction curve is the final result and is displayed as ng/ml. The upper limit of the normal range was 120 ng/ml.

To avoid any bias, the haematology staff were unaware of the venogram results. The radiology staff performing the venograms were also blinded to the D-dimer results. The emergency department performed the Simplify D-dimer test before the venogram.

Table 2 Sensitivity, specificity, NPV, and PPV of the investigated D-dimer assays

Test	Sensitivity %	Specificity %	NPV %	PPV %
Simplify	94.1 (87.7 to 100)	40.4 (32.2 to 48.7)	94.8 (91.0 to 98.6)	37.2 (28.9 to 45.6)
SimpliRED	74.5 (62.5 to 86.5)	83.1 (76.8 to 89.4)	89.7 (82.0 to 97.3)	62.3 (50.1 to 74.5)
Auto D-dimer	90.2 (82.0 to 98.4)	43.4 (35.1 to 51.7)	92.2 (87.4 to 96.9)	37.4 (28.8 to 45.9)

The figures in parentheses are the 95% confidence intervals.

The sensitivity, specificity, positive predictive value (PPV), NPV, and 95% confidence intervals (CI) were calculated.

RESULTS

A total of 187 patients were enrolled into this study, 86 men and 101 women. Table 1 summarises the results and table 2 lists the sensitivity, specificity, NPV, and PPV for each of the D-dimer tests.

DISCUSSION

In this study 51 (27.2%) patients had a positive venogram and the Simplify D-dimer test proved to have the highest sensitivity of the three methods (table 2). The Simplify test "missed" three cases of DVT. In one of the cases (a 66 year old woman) with calf vein DVT and a false negative Simplify test both the latex agglutination and SimpliRED tests were also negative. Two male patients (aged 39 and 80) with proximal DVT also had a false negative Simplify test. In both these cases the SimpliRED and latex agglutination tests were both positive. The latex agglutination test had five false negative results and the SimpliRED test 13 false negative results.

With any new D-dimer test it is mandatory to have a high sensitivity and to have few false negative D-dimer results. With the SimpliRED test there is a visual end point of agglutination of red blood cells. It is much easier to interpret the Simplify test as the positive end point is a pink/purple line. It is expected that the interobserver variability will be less with the Simplify test than the SimpliRED test.

With quantitative D-dimer assays it is important to select a cut off level such that the D-dimer assay has a high sensitivity. However, the cut off level should not be so low that the specificity of the assay is also very low. A cut off level must be chosen such that there is a balance between sensitivity and specificity. The upper limit of the normal D-dimer range was initially chosen to be 120 ng/ml for the latex agglutination test (based on a normal cohort of laboratory staff). The five false negative D-dimer results were 70 ng/ml or less. Unless the cut off level was changed to less than 70 ng/ml the sensitivity of the assay would not be improved and there would be a very low specificity at a cut off level of 70 ng/ml.

It has been reported that there is a significantly lower sensitivity for the SimpliRED test if citrated whole blood is used instead of fingerstick capillary samples.¹⁵ Although the manufacturers of SimpliRED suggest that fingerstick capillary blood can be used, in the emergency department it is usually easier to take venous blood, as other blood investigations often need to be performed. In this study we used citrated venous blood for the SimpliRED test because the sample was sent to the haematology laboratory. The Simplify test was performed on venous blood.

The increased sensitivity of the Simplify D-dimer test when compared with the other two tests was at the expense of specificity (table 2). The specificity of the Simplify test was only 40.4%, compared with SimpliRED (83.1%) and latex agglutination (43.4%).

CONCLUSION

The Simplify D-dimer is easy to use, can be performed in 10 minutes, and the sensitivity was 94.1% in this study. This test is more suitable for use in the emergency department than the SimpliRED test. However, despite the high NPV of 94.8%, this test should be used in conjunction with clinical probability scoring, plethysmography, or ultrasonography to reach failure rates of less than 1%.

The Simplify test costs about £7, which is similar to most D-dimer tests. However, the laboratory based tests also need an analyser that entails a considerable initial capital cost. The cost of contrast for venography is on average over £30 and more importantly, ultrasonography and venography are very time consuming for skilled hospital staff. Despite the low specificity of the Simplify test, any reduction in further imaging would make the test worthwhile from an economic viewpoint.

CONTRIBUTORS

DN drafted the original manuscript, did the statistical analysis, and the laboratory investigations. CT initiated and coordinated the study, examined the subjects, and did the Simplify tests. AV was the radiology lead for the study and performed many of the venograms. SD did many of the laboratory investigations. KM was the haematology lead for the study. MO, VR, and AH examined the subjects and did the Simplify tests. All authors contributed to the final draft of the paper. CT is the guarantor.

A colour version of the figure is available on the journal web site (<http://www.emjonline.com/supplemental>).

Authors' affiliations

D Neale, S Davies, K Myers, Haematology Department, Prince Charles Hospital, Merthyr Tydfil, Mid Glamorgan, Wales

C Tovey, M Obiako, V Ramkumar, A Hafiz, Emergency Department, Prince Charles Hospital

A Vali, Radiology Department, Prince Charles Hospital

Funding: none.

Competing interests: none declared.

REFERENCES

- 1 **Carter CJ**. Epidemiology of venous thromboembolism. In: Hull R, Pineo GF, eds. *Disorders of thrombosis*. Philadelphia: WB Saunders, 1996:159-74.
- 2 **Ebell MH**. Evaluation of the patient with suspected deep vein thrombosis. *J Fam Pract* 2001;**50**:167-71.
- 3 **O'Donnell TF Jr**, Abbott WM, Athanasoulis CA, et al. *Surg Gynecol Obstetr* 1980;**150**:69-74.
- 4 **deValois JC**, van Schaik CC, Verzijlbergen F, et al. Contrast venography: from gold standard to 'golden backup' in clinically suspected deep vein thrombosis. *Eur J Radiol* 1990;**11**:131-7.
- 5 **Heijboer H**, Büller HR, Lensing AWA, et al. A comparison of real-time compression ultrasonography with impedance plethysmography for the diagnosis of deep-vein thrombosis in symptomatic outpatients. *N Engl J Med* 1993;**329**:1365-9.
- 6 **Cogo A**, Lensing AWA, Koopman MMW, et al. Compression ultrasonography for diagnostic management of patients with clinically suspected deep vein thrombosis: prospective cohort study. *BMJ* 1998;**316**:17-20.
- 7 **Tovey C**, Wyatt S. Diagnosis, investigation, and management of deep vein thrombosis. *BMJ* 2003;**326**:1180-4.

- 8 Kelly J, Rudd A, Lewis RR, *et al*. Plasma D-dimers in the diagnosis of venous thromboembolism. *Arch Intern Med* 2002;**162**:747–56.
- 9 Schutgens REG, Haas FJLM, Gerritsen WBM, *et al*. The usefulness of five D-dimer assays in the exclusion of deep venous thrombosis. *Journal of Thrombosis and Haemostasis* 2003;**1**:976–81.
- 10 Dempfle CE, Zips S, Ergül H, *et al*. The fibrin assay comparison trial (FACT): evaluation of 23 quantitative D-dimer assays as basis for the development of D-dimer calibrators. *Thromb Haemost* 2001;**85**:671–8.
- 11 Lennox AF, Delis KT, Serunkuma S, *et al*. Combination of a clinical risk assessment score and rapid whole blood D-dimer testing in the diagnosis of deep vein thrombosis in symptomatic patients. *J Vasc Surg* 1999;**30**:794–803.
- 12 Chapman CS, Akhtar N, Campbell S, *et al*. The use of D-dimer assay by enzyme immunoassay and latex agglutination techniques in the diagnosis of deep vein thrombosis. *Clin Lab Haematol* 1990;**12**:37–42.
- 13 Eisenberg PR. Does a negative D-dimer exclude thrombosis? *Fibrinolysis* 1993;**7**(suppl 2):32–5.
- 14 Turkstra F, van Beek EJR, Büller HR. Observer and biological variation of a rapid whole blood D-dimer test. *Thromb Haemost* 1998;**79**:91–3.
- 15 Mauron T, Baumgartner I, Z'Brun A, *et al*. SimpliRED D-dimer assay: comparability of capillary and citrated venous whole blood, between-assay variability, and performance of the test for exclusion of deep vein thrombosis in symptomatic outpatients. *Thromb Haemost* 1998;**79**:1217–19.
- 16 American Thoracic Society. The diagnostic approach to acute venous thromboembolism. *Am J Respir Crit Care Med* 1999;**160**:1043–66.
- 17 Kraaijenhagen RA, Lensing AWA, Wallis JW, *et al*. Diagnostic management of venous thromboembolism. *Ballieres Clin Haematol* 1998;**11**:541–85.
- 18 Aschwanden M, Labs KH, Jeanneret C, *et al*. The value of rapid D-dimer testing combined with structured clinical evaluation for the diagnosis of deep vein thrombosis. *J Vasc Surg* 1999;**30**:929–35.
- 19 Kearon C, Ginsberg JS, Douketis J, *et al*. Management of suspected deep venous thrombosis in outpatients by using clinical assessment and D-dimer testing. *Ann Intern Med* 2001;**135**:108–11.
- 20 Ginsberg JS, Kearon C, Douketis J, *et al*. The use of D-dimer testing and impedance plethysmographic examination in patients with clinical indications of deep vein thrombosis. *Arch Intern Med* 1997;**157**:1077–81.
- 21 van der Graaf F, van den Borne H, van der Kolk M, *et al*. Exclusion of deep vein thrombosis with D-dimer testing. Comparison of 13 D-dimer methods in 99 outpatients suspected of deep venous thrombosis using venography as reference standard. *Thromb Haemost* 2000;**83**:191–8.
- 22 John MA, Elms MJ, O'Reilly EJ, *et al*. The SimpliRED D-dimer test: a novel assay for the detection of cross-linked fibrin degradation products in whole blood. *Thromb Res* 1990;**58**:273–81.

CONTINUING MEDICAL EDUCATION.....

ACCESS: the acute cerebral CT evaluation stroke study

Do you interpret CT brain scans within the first few hours of stroke? We run the acute cerebral CT evaluation stroke study (ACCESS), an internet based, interactive, CT reading tool designed to evaluate and improve CT reader reliability in detecting early infarct sign on CT. Readers log on to a web server (<http://www.neuroimage.co.uk>), complete a few details about background training and experience in viewing CT scans, view the study CT scans, and answer questions about each scan on the same screen. It is our intention to investigate the cross disciplinary recognition of stroke and to improve upon pre-existing paradigms in the diagnosis of acute stroke on CT.

There are 56 scans, reviewed in batches of about 10 scans each of which take about 20 minutes to complete. We will present the results of the study once it is completed later this year. Participation in ACCESS counts towards five continuing medical education category 1 credits in the UK. Accreditation was awarded by the Royal College of Radiologists, but because accreditation is not specific to RCR members, there is cross recognition between all UK Royal Colleges.

J Wardlaw

University of Edinburgh, Department of Clinical Neurosciences, Western General Hospital, Crewe Road, Edinburgh EH4 2XU, UK; jmw@skull.dcn.ed.ac.uk