

## LETTER

## Use of negative pressure individual isolation system for CT scan of patients with suspected COVID-19

Dear Editor

The Sabah Women's and Children's Hospital in Malaysia has one CT scan. Using the single scanner to image patients with suspected COVID-19 would cause a backlog, due to the need for frequent and thorough disinfection after each case. In the care of recent COVID-19 suspect, with a negative CXR, we carried out a CT thorax using a negative pressure individual isolation system (NPIIS), Iso-Pod<sup>1</sup> with our Siemens SOMATOM Definition AS make<sup>2</sup> CT scanner and describe our experience.

The NPIIS was laid out on a transfer stretcher and the patient lay down on top of it. The NPIIS was then assembled around her by the accompanying physician. This process took less than 5 min. The NPIIS is predominantly plastic with minimal metal materials in its structure. The size is suitable to fit into the CT scan gantry comfortably (figure 1). The patient



**Figure 1** Negative pressure individual isolation system fitting comfortably into the CT scan gantry.

in the NPIIS was taken to CT, which was carried out using similar settings as when a CT thorax is done without the NPIIS. The spatial resolution of this scan was found to be equivalent and no noticeable noise on the image quality was seen. We observed minimal metal streak artefacts from the metallic material in the NPIIS which was outside of the patient's body. These artefacts did not interfere with the CT thorax interpretation and image quality was excellent for diagnostic purposes. After use, the NPIIS was wiped with a dish-washing sponge soaked in detergent to remove visible soiling and stains. It was then soaked in 0.1% sodium hypochlorite solution for 1 hour and rinsed with water, then air dried for reuse, following our hospital infectious disease prevention protocol for usage of NPIIS.

Using an NPIIS may enable the work process of conducting a CT for patients with suspected COVID-19 to be less disruptive to the hospital operations and would limit needs for disinfection while reducing the use of personal protective equipment (PPE). In conducting this case, only one clinician was designated to be in full PPE on standby for emergency resuscitation. However, our patient was fairly stable.

For critically ill patients, the process of transferring the patient into the NPIIS may cause delay and is a risk in itself; therefore, conducting a CT scan without using NPIIS may be a better option for these individuals. The use of an NPIIS should be governed by local institutional protocols, suitability of patients and adherence to safety precautions.

**Phee Kheng Cheah** <sup>1</sup>, **Thiagu Krisnan**,<sup>2</sup>  
**Mohamed Hakimi Abdul Kadir**,<sup>1</sup>  
**Evelyn Marie Steven**<sup>1</sup>

<sup>1</sup>Emergency and Trauma Department, Sabah Women and Children's Hospital, Ministry of Health Malaysia, Kota Kinabalu, Sabah, Malaysia

<sup>2</sup>Department of Radiology, Sabah Women and Children's Hospital, Ministry of Health Malaysia, Kota Kinabalu, Malaysia

**Correspondence to** Dr Phee Kheng Cheah, Emergency and Trauma Department, Sabah Women and Children's Hospital, Ministry of Health Malaysia, 88996 Kota Kinabalu, Sabah, Malaysia; dr.cpkheng@gmail.com

**Acknowledgements** The authors would like to thank the Director General of Health Malaysia for permission to publish this letter.

**Contributors** All authors contributed equally in the conception, production and review of this manuscript.

**Funding** The authors have not declared a specific grant for this research from any funding agency in the public, commercial or not-for-profit sectors.

**Competing interests** None declared.

**Patient and public involvement** Patients and/or the public were not involved in the design, or conduct, or reporting, or dissemination plans of this research.

**Patient consent for publication** Obtained.

**Provenance and peer review** Not commissioned; internally peer reviewed.

This article is made freely available for use in accordance with BMJ's website terms and conditions for the duration of the covid-19 pandemic or until otherwise determined by BMJ. You may use, download and print the article for any lawful, non-commercial purpose (including text and data mining) provided that all copyright notices and trade marks are retained.

© Author(s) (or their employer(s)) 2020. No commercial re-use. See rights and permissions. Published by BMJ.

**Handling editor** Lara Nicole Goldstein



**To cite** Cheah PK, Krisnan T, Abdul Kadir MH, *et al*. *Emerg Med J* 2020;**37**:467.

Accepted 4 June 2020  
Published Online First 16 June 2020

*Emerg Med J* 2020;**37**:467. doi:10.1136/emered-2020-209851

### ORCID iD

Phee Kheng Cheah <http://orcid.org/0000-0002-2523-2997>

### REFERENCES

- 1 Anonymous. Airboss of America Corp. Available: <https://airbossofamerica.com/engineered-products/defense/products/cbrn/isopod-chem-bio.php#> [Accessed 21 Apr 2020].
- 2 Anonymous. Siemens medical solutions. Available: <https://www.siemens-healthineers.com/en-us/computed-tomography/ecoline-refurbished-systems/somatomdefinitionas> [Accessed 19 May 2020].